

Radiology Compliance Branch
RADIATION PROTECTION SECTION



Division of Health Service Regulation • NC Department of Health and Human Services

REGISTRANTS MOST FREQUENT VIOLATIONS

1. The facility Written Radiation Protection program was not available or not adequate.
2. Not maintaining the appropriate records (Notice of Registration (NOR), Plan Review, Report of Assembly (FDA 2579), Letter of Acknowledgement (LOA), Post room survey, and annual review of Written Safety program).
3. Personnel monitoring equipment has not been supplied and/or used by all occupationally exposed personnel.
4. The registrant failed to update the information contained in the application for registration or notice of registration no longer accurate.
5. The registrant failed to annually review the Written Radiation Protection program.
6. The registrant failed to have a copy of the "North Carolina Regulations For Protection Against Radiation" at the facility.
7. The registrant failed to provide a working technique chart for each diagnostic X-ray system.
8. There has not been an area radiation survey done within 30 days following initial operation of equipment.
9. Service Provider failed to submit records showing receipt, transfer, and/or disposal of radiation sources.
10. Annual registration fees not submitted.

